



# Multicenter evaluation of the Vetscan Imagyst system using Ocus 40 and EasyScan One scanners to detect gastrointestinal parasites in feces of dogs and cats

Yoko Nagamori,<sup>1</sup> Ruth Scimeca, Ruth Hall-Sedlak, Byron Blagburn, Lindsay A. Starkey, Dwight D. Bowman, Araceli Lucio-Forster, Susan E. Little, Travis Cree, Michael Loenser, Benjamin S. Larson, Cory Penn, Austin Rhodes, Richard Goldstein

**Abstract.** The Vetscan Imagyst system (Zoetis) is a novel, artificial intelligence–driven detection tool that can assist veterinarians in the identification of enteric parasites in dogs and cats. This system consists of a sample preparation device, an automated digital microscope scanner, and a deep-learning algorithm. The EasyScan One scanner (Motic) has had good diagnostic performance compared with manual examinations by experts; however, there are drawbacks when used in veterinary practices in which space for equipment is often limited. To improve the usability of this system, we evaluated an additional scanner, the Ocus 40 (Grundium). Our objectives were to 1) qualitatively evaluate the performance of the Vetscan Imagyst system with the Ocus 40 scanner for identifying *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, *Cystoisospora* oocysts, and *Giardia* cysts in canine and feline fecal samples, and 2) expand the assessment of the performance of the Vetscan Imagyst system paired with either the Ocus 40 or EasyScan One scanner to include a larger dataset of 2,191 fecal samples obtained from 4 geographic regions of the United States. When tested with 852 canine and feline fecal samples collected from different geographic regions, the performance of the Vetscan Imagyst system combined with the Ocus 40 scanner was correlated closely with manual evaluations by experts. Sensitivities were 80.0–97.0% and specificities were 93.7–100.0% across the targeted parasites. When tested with 1,339 fecal samples, the Vetscan Imagyst system paired with the EasyScan One scanner successfully identified the targeted parasite stages; sensitivities were 73.6–96.4% and specificities were 79.7–100.0%.

**Keywords:** artificial intelligence; cats; detection; dogs; intestinal parasites; parasitology.

Fecal examination for the detection of gastrointestinal parasitism in dogs and cats is an important component of routine veterinary care.<sup>7</sup> Although various techniques can be used to examine fecal samples for evidence of parasitic infection, centrifugal or passive fecal flotation followed by microscopic examination for parasite stages, such as eggs, oocysts, and cysts, is a standard practice in most veterinary clinics. However, the precision of these traditional fecal examinations can vary widely and be affected by procedural differences and experience levels of the personnel conducting the tests, possibly leading to poor recovery and errors of identification.<sup>1,7,12,22</sup> To improve the accuracy, reliability, and consistency of fecal examinations in veterinary clinical settings, innovative computer technology–driven algorithms have been developed.<sup>8,17,26,27,33,34</sup> Among these developments, the Vetscan Imagyst (Zoetis) is an artificial intelligence (AI)-based system that can be utilized in veterinary clinics to detect eggs, oocysts, and cysts of common parasites in canine and feline fecal samples. The Vetscan Imagyst enables

veterinary clinics to provide a consistent fecal examination method that is not affected by an examiner’s level of experience.<sup>26,27,41</sup>

The Vetscan Imagyst system has a consistent ability to detect and correctly identify the eggs of *Ancylostoma*, *Toxocara*, *Trichuris*, and taeniid tapeworms, as well as *Cystoisospora* oocysts and *Giardia* cysts in fecal samples from dogs and cats.<sup>26,27</sup> The system consists of 3 key components: a

Zoetis, Global Diagnostics, Parsippany, NJ, USA (Nagamori, Hall-Sedlak, Cree, Loenser, Penn, Rhodes, Goldstein); Oklahoma Animal Disease Diagnostic Laboratory (Scimeca) and Department of Veterinary Pathobiology (Little), College of Veterinary Medicine, Oklahoma State University, Stillwater, OK, USA; Department of Pathobiology, College of Veterinary Medicine, Auburn University, Auburn, AL, USA (Blagburn, Starkey); Department of Microbiology and Immunology, College of Veterinary Medicine, Cornell University, Ithaca, NY, USA (Bowman, Lucio-Forster); Techeyte, Lindon, UT, USA (Larson).

<sup>1</sup>Corresponding author: Yoko Nagamori, Zoetis, Global Diagnostics, 10 Sylvan Way, Parsippany, NJ 07054, USA. yoko.nagamori@zoetis.com

**Table 1.** Technical specifications of the automated microscope scanners (Ocus 40, Grundium; EasyScan One, Motic) used with the Vetscan Imagyst (Zoetis).<sup>14,25</sup>

	Ocus 40	EasyScan One
Objective lens	0.75 na 20× Plan Apo	0.75 na 20× Plan Apo
Accommodates multiple objective lenses	No	Yes
Effective magnification	0.25 mpp	0.25 mpp
Scan time (20 × 20 mm)	8–10 min	4–6 min
Image quality	Excellent	Excellent
Lighting and image capture	Blue, green, and red LED strobe with a rolling shutter grayscale sensor	White LED illumination with multicolor sensor
Scanner dimensions and computer control	18 × 18 × 19 cm; scanner with integrated GPU, allowing user interface from basic internet connected device (Fig. 1)	20 × 40 × 42 cm; connected physically (wired) to a control laptop with GPU capable of handling images (Fig. 2)
Loading and cleaning	Easier front-loading system; objective lens accessible for inspection/cleaning (Fig. 1)	More complicated slide tray loading; cover must be removed to view/clean objective lens (Fig. 2).
User interface	Combined scanner control and interface for viewing results	Separate user interface for scanner control and viewing results

GPU = graphic processing unit; LED = light emitting diode; mpp =  $\mu\text{m}$  per pixel; na = numerical aperture; Plan Apo = plan apochromat.

sample preparation device, an automated digital microscope scanner, and an AI-based data analytical algorithm using a deep-learning convolutional neural network (CNN). Studies with the Vetscan Imagyst system have shown high sensitivity (75.8–100.0%) and specificity (93.1–100.0%) across 6 targeted parasites compared to conventional centrifugal fecal flotation examinations performed by parasitologists.<sup>26,27</sup> The capabilities of the Vetscan Imagyst system can be improved and expanded for additional parasites with supplemental training, as mentioned in previous studies.<sup>26,27</sup>

The EasyScan One scanner (Motic) has exhibited high image quality and fast scanning time, resulting in the successful identification of various forms of targeted parasites.<sup>26,27</sup> However, the relatively large physical dimensions of the EasyScan One scanner, together with the requirement for a physical connection to a computer, may be problematic at many veterinary clinics in which physical space for laboratory equipment is at a premium. Additionally, the requirement for a slide tray for loading the prepared fecal slide into the scanner was considered cumbersome by some users. To address these concerns, we sought an alternative scanner based on smaller physical dimensions, high image quality, optics, and practicality and usability in the veterinary clinical setting. The Ocus 40 scanner (Grundium) met these specifications and was thus selected for further evaluation with the Vetscan Imagyst system.

In studies of Vetscan Imagyst performance, 188 canine and 112 feline fecal samples had been examined.<sup>26,27</sup> Given that the collection of these samples was limited to central Oklahoma, the potential impact of variations in fecal components in different geographic regions was not addressed, even though certain gastrointestinal parasites, as well as microscopic plant pollens and seeds that may resemble

parasite eggs, oocysts, and cysts, can vary across geographic regions.<sup>23,35</sup> Thus, one of our aims was to overcome this limitation by extending the fecal sampling to other geographic regions. We tested the Vetscan Imagyst system using a large number of canine and feline fecal samples collected from 4 different geographic regions in the United States.

Our study objectives were 1) to qualitatively evaluate the performance of the Vetscan Imagyst system with 2 commercial scanners—Ocus 40 and EasyScan One—for correctly identifying *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, *Cystoisospora* oocysts, and *Giardia* cysts in feces of dogs and cats, and 2) to assess the performance of the Vetscan Imagyst system among multiple users, using a large number of fecal samples collected from 4 different geographic regions.

## Materials and methods

### Technical specifications of Ocus 40 and EasyScan One scanners

The Ocus 40 and EasyScan One scanners capture images with an effective magnification of 0.25  $\mu\text{m}$  per pixel (mpp), have similar objective lenses (0.75 numerical aperture, 20× plan apochromat) and comparable scanning times for a 20 × 20 mm area, and provide high-quality images (Table 1). The Ocus 40 scanner uses a multi-color strobe of light-emitting diodes (LEDs), wherein several images are captured with blue, green, and red LED flashes; these images are assembled into a single image for analysis.<sup>31</sup> In contrast, the EasyScan One scanner uses a single 10-watt white LED light and captures a single multicolor image.<sup>25</sup> The Ocus 40 scanner is more compact (18 × 18 × 19 cm; Fig. 1)<sup>14</sup> than the EasyScan One scanner (20 × 40 × 42 cm; Fig. 2).<sup>25</sup> Additionally,



**Figure 1.** The Ocus 40 scanner (Grundium) has dimensions of  $18 \times 18 \times 19$  cm, and contains an integrated graphic processing unit, allowing a user interface with a basic internet-connected device (laptop shown for scale).



**Figure 2.** The EasyScan One scanner (Motic) has dimensions of  $20 \times 40 \times 42$  cm, and must be connected physically (wired) to a control laptop with a graphic processing unit capable of handling images.

the Ocus 40 scanner has a simpler front-loading system, allowing users to place a slide directly on the stage and inspect and clean the objective lens quickly and easily (Fig. 1). In contrast, the EasyScan One scanner requires a slide tray to load a sample slide, and the cover that encloses the objective lens must be removed to allow inspection and cleaning (Fig. 2).

To qualify for integration with the Vetscan Imagyst system, a scanner must be compatible, allowing the Vetscan Imagyst system to configure settings and control the scanner through an application programming interface (API). Assessment with the Vetscan Imagyst system has shown the Easy Scan One scanner to be moderately configurable, allowing partial control with the API. In contrast, the Ocus 40 scanner was highly configurable, allowing a higher level of control with the API. The Ocus 40 scanner also includes an integrated graphic processing unit (GPU) and a network

interface, allowing it to be used as a standalone device that users can interface with and control from any computer with an internet connection. The EasyScan One scanner, on the other hand, must be physically connected to a computer containing a GPU that is capable of rapid image capture.

Subjectively, the quality of images produced by both scanners is excellent: clear, sharp, well-focused, high resolution, low noise, and limited distortion. However, there are noticeable color distinctions in images generated by the Ocus 40 and EasyScan One scanners. Factors that can influence image colors include saturation, intensity, and contrast; different post-processing corrections can influence these color variations.<sup>4,10,11</sup> The Ocus 40 scanner employs a multi-color flash of LEDs in which the individual flashes are captured using a rolling shutter grayscale sensor.<sup>36</sup> To get an accurate color presentation of the combined grayscale images that represent captured color channels, the Ocus 40 scanner

utilizes a patented color correction method,<sup>30</sup> which results in more image information per pixel with greater depth and detail for the same resolution sensor. Additionally, the Ocus 40 scanner uses a unique image-stitching technique; at least 3 source images are composited to make an image while proposing a candidate transformation for each image pair and solving edge weights with an optimization algorithm to determine plausibility of the transformations. Generally, this procedure is thought to provide better quality images compared to techniques that have only one image to leverage for matching features and edges to stitch individual adjacent fields of view.<sup>29</sup>

### Fecal sample collections, pre-screening, and labeling

Fecal sample collection and the Vetscan Imagyst evaluation were conducted from August 2019 to September 2021 at 3 centers: Oklahoma Animal Disease Diagnostic Laboratory (OADDL; Oklahoma State University, Stillwater, OK, USA); Auburn University–College of Veterinary Medicine (AU-CVM; Auburn, AL, USA); and Cornell University–College of Veterinary Medicine (CU-CVM; Ithaca, NY, USA). Additional fecal sample collection was coordinated at the Zoetis Reference Laboratory (Mukilteo, WA, USA), and fecal samples collected at this location were mailed to and assessed at the 3 centers. Most fecal samples originated locally from OK, AL, NY, and WA. Our study did not require regulatory review or approval by Institutional Animal Care and Use Committees because animals were not handled either directly or indirectly.

Fecal samples submitted from dogs and cats undergoing routine fecal examinations were processed using the modified Wisconsin fecal examination technique with sugar solution (specific gravity: 1.25–1.27) or zinc sulfate solution (specific gravity: 1.18),<sup>39</sup> without prior randomization. Fecal samples ( $\geq 1$  g) containing any targeted parasites (*Ancylostoma*, *Cystoisospora*, *Giardia*, *Toxocara*, and/or *Trichuris*) and fecal samples ( $\geq 5$  g) that were not observed to contain any parasites were included in the study. All fecal samples were stored at 4°C and examined within 14 d of the submittal date.

### Assessment of Vetscan Imagyst system performance

The EasyScan One scanner was utilized in previous studies and from August 2019 to June 2020 in our study; we evaluated the Ocus 40 scanner from July 2020 to September 2021. A deep-learning CNN was applied to the Vetscan Imagyst system as described previously.<sup>26,27</sup> The EasyScan One scanner was combined with the algorithm v.3033, which we also utilized in our previous study and had been largely developed and trained using images generated by the EasyScan One scanner.<sup>27</sup> For the Ocus 40 scanner, we used a new version of the algorithm, v.8293, which was specifically

designed and adapted to the characteristics of images produced by the Ocus 40 scanner.

Fecal examination slides were prepared using the Vetscan Imagyst sample preparation device according to the instructions described previously<sup>26,27</sup> or in the Vetscan Imagyst Quick Start Guide.<sup>42</sup> Samples evaluated for *Giardia* were prepared in zinc sulfate solution (specific gravity: 1.18–1.20), and the rest of the samples were prepared in sugar solution (specific gravity: 1.28–1.30; Zoetis data on file). Slides were then analyzed by the Vetscan Imagyst system and manually by experienced laboratory personnel. Fecal slide examination results were compared, and the sensitivity and specificity of the Vetscan Imagyst algorithm paired with the Ocus 40 and EasyScan One scanners were calculated.

### Statistical analysis

A fecal sample was scored as “positive” if one or more of the targeted parasites were detected by an expert veterinary parasitologist (OADDL: R. Scimeca, S.E. Little, Y. Nagamori; AU-CVM: B. Blagburn, L.A. Starkey; CU-CVM: D.D. Bowman, A. Lucio-Forster). When multiple targeted parasites were present in a sample, samples were counted as positive for more than one analysis; some samples were counted as negative for more than one analysis. Using positive and negative determinations from both the expert and Vetscan Imagyst (negative corresponding to a count of 0, positive corresponding to counts  $\geq 1$ ),  $2 \times 2$  contingency tables were constructed for each targeted parasite by scanner and sample preparation types. Using expert determination as the gold standard, sensitivity, specificity, and 95% Clopper–Pearson exact CIs were estimated from the proportions of samples, determined to be true-positive (TP) and true-negative (TN) among the positive (TP+false-negative [FN]) and negative (TN+false-positive [FP]) samples, respectively. In addition to Clopper–Pearson exact CIs, Jeffrey 95% CIs were constructed. R v.4.2.1 (<https://www.r-project.org/>) was used for the statistical analyses.

### Results

We examined 2,191 fecal samples for the detection of canine and feline *Ancylostoma*, *Toxocara*, *Trichuris* eggs, *Cystoisospora* oocysts, and *Giardia* cysts. Of the 2,191 fecal samples, 852 were evaluated by the Ocus 40 scanner paired with the v.8293 algorithm, and 1,339 were evaluated by the EasyScan One scanner paired with the v.3033 algorithm (Table 2).

For *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, and *Cystoisospora* oocysts in canine fecal samples, the diagnostic sensitivity and specificity (DSe and DSp, respectively) of the Vetscan Imagyst system with the Ocus 40 scanner compared with the experts’ assessments were 91.2–97.0% and 96.7–99.7%, respectively (Table 3). The DSe and DSp of the Vetscan Imagyst system with the EasyScan One scanner

**Table 2.** Summary of expert characterized fecal samples by type of scanner (Ocus 40, EasyScan One) and parasite detection.

	Scanner, flotation solution	No. of samples	Parasites detected	No. of samples
Canine samples	Ocus 40, sugar solution	502	No parasite	92
			1 parasite	269
			>1 parasite	141
	EasyScan One, sugar solution	943	No parasite	286
			1 parasite	476
			>1 parasite	181
Feline samples	Ocus 40, sugar solution	203	No parasite	67
			1 parasite	104
			>1 parasite	32
	EasyScan One, sugar solution	309	No parasite	105
			1 parasite	169
			>1 parasite	35
<i>Giardia</i> samples; canine and feline	Ocus 40, zinc sulfate solution	147	No parasite	84
			<i>Giardia</i> detected	63
	EasyScan One, zinc sulfate solution	87	No parasite	15
			<i>Giardia</i> detected	72

**Table 3.** Algorithm performance with the Ocus 40 scanner for canine fecal samples prepared with the Vetscan Imagyst sample preparation device with sugar solution: diagnostic sensitivity and specificity comparing the results reported by an expert (reference) versus by the Vetscan Imagyst system.

	Ancylostoma	Cystoisospora	Toxocara	Trichuris
True-positive	197	78	97	122
True-negative	279	407	396	370
False-positive	7	14	6	1
False-negative	19	3	3	9
Total	502	502	502	502
Sensitivity, % (95% CI)	91.2 (86.9–94.4)	96.3 (90.4–98.9)	97.0 (92.2–99.1)	93.1 (87.8–96.5)
Specificity, % (95% CI)	97.6 (95.3–98.9)	96.7 (94.6–98.1)	98.5 (96.9–99.4)	99.7 (98.7–100.0)

compared with the experts' assessments were 89.1–96.3% and 79.7–96.0%, respectively (Table 4).

For the same targeted parasites in feline fecal samples, the DSe and DSp of the Vetscan Imagyst system with the Ocus 40 scanner compared with the experts' assessments were 80.0–93.5% and 93.7–100.0%, respectively (Table 5). With the EasyScan One scanner, the DSe and DSp of the Vetscan Imagyst system compared with the experts' assessments were 91.4–96.4% and 92.7–98.2%, respectively (Table 6).

For the *Giardia* analysis, results of canine and feline fecal assessments were combined because of a recording error. The DSe and DSp of the Vetscan Imagyst system with the Ocus 40 scanner for *Giardia* detection were 92.1% and 98.8%, respectively. With the EasyScan One scanner, the DSe and DSp of the Vetscan Imagyst for *Giardia* cysts were estimated as 73.6% and 100.0%, respectively (Table 7).

Although the number of FP results decreased with the Ocus 40 scanner combined with the v.8293 algorithm, the most frequently observed FP result throughout the study was *Eimeria* oocysts misclassified as *Cystoisospora* oocysts.

Additionally, during the evaluation of the EasyScan One scanner combined with the v.3033 algorithm, some other FP results were noticed: yellow debris confused with *Trichuris* eggs, end-on or tilted *Ancylostoma* eggs misclassified as *Cystoisospora* oocysts, and various plant materials misidentified as eggs of *Ancylostoma*, *Toxocara*, and *Trichuris*, as well as *Cystoisospora* oocysts.

Occasionally the Vetscan Imagyst system detected and classified targeted parasites successfully even when the human eye missed them at the initial manual reading. Specifically, the Vetscan Imagyst system correctly reported 2 cases of *Ancylostoma* eggs, 11 cases of *Cystoisospora* oocysts, 3 cases of *Giardia* cysts, 3 cases of *Toxocara* eggs, and 1 case of *Trichuris* eggs that had evaded human detection during their examinations. For most of these cases, there were very few parasite forms, generally only 1 or 2 eggs, oocysts, or cysts, observed on the entire slide. In these cases, a boarded diagnostic parasitologist (Y. Nagamori) re-reviewed the whole slide images to determine if these cases were TPs or FPs.

**Table 4.** Algorithm performance with the EasyScan One scanner for canine fecal samples prepared with the Vetscan Imagyst sample preparation device with sugar solution: diagnostic sensitivity and specificity comparing the results reported by an expert (reference) versus by the Vetscan Imagyst system.

	Ancylostoma	Cystoisospora	Toxocara	Trichuris
True-positive	284	167	209	114
True-negative	619	606	694	762
False-positive	29	154	29	53
False-negative	11	16	11	14
Total	943	943	943	943
Sensitivity, % (95% CI)	96.3 (93.6–98.0)	91.3 (86.5–94.7)	95.0 (91.5–97.3)	89.1 (82.8–93.6)
Specificity, % (95% CI)	95.5 (93.7–96.9)	79.7 (76.8–82.5)	96.0 (94.4–97.2)	93.5 (91.6–95.0)

**Table 5.** Algorithm performance with the Ocus 40 scanner for feline fecal samples prepared with the Vetscan Imagyst sample preparation device with sugar solution: diagnostic sensitivity and specificity comparing the results reported by an expert (reference) versus by the Vetscan Imagyst system.

	Ancylostoma	Cystoisospora	Toxocara	Trichuris
True-positive	8	71	72	0
True-negative	188	119	123	203
False-positive	5	8	3	0
False-negative	2	5	5	0
Total	203	203	203	203
Sensitivity, % (95% CI)	80.0 (49.7–95.6)	93.4 (86.2–97.4)	93.5 (86.4–97.5)	NA
Specificity, % (95% CI)	97.4 (94.4–99.0)	93.7 (88.5–97.0)	97.6 (93.8–99.3)	100.0 (98.8–100.0)

One *Ancylostoma* fecal sample was excluded from analysis because the sample was not examined by the algorithm because of an error.  
NA=not applicable.

## Discussion

The pairing of the Ocus 40 microscope scanner with the Vetscan Imagyst system resulted in high sensitivities and specificities for *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, *Cystoisospora* oocysts, and *Giardia* cysts in feces of dogs and cats. DSe and DSp observed for targeted parasites in canine fecal samples were comparable between the Ocus 40 and EasyScan One scanners, with mostly overlapping 95% CIs. Specificities for *Cystoisospora* oocysts (96.7%; 95% CI: 94.6–98.1%) and *Trichuris* eggs (99.7%; 95% CI: 98.7–100.0%) recorded with the Ocus 40 scanner were higher than specificities for *Cystoisospora* oocysts (79.7%; 95% CI: 76.8–82.5%) and *Trichuris* eggs (93.5%; 95% CI: 91.6–95.0%) recorded with the EasyScan One scanner, with no overlap of 95% CIs. These increases in specificity may be explained by differences in the development and presentation of colors in images and the process of generating images by the Ocus 40 scanner given that the color variations can be pivotal for the deep-learning object detection algorithm because its performance is affected by colors and other details, including texture, edges, shape, and size of the objects.<sup>4,10,11</sup> Additionally, the new algorithm, v.8293, reconciled well with the Ocus 40 scanner, and the good compatibility of the algorithm and scanner may have enabled the Vetscan Imagyst system to identify *Cystoisospora* oocysts and *Trichuris* eggs more accurately.

In feline fecal samples, diagnostic sensitivities and specificities to detect *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, and *Cystoisospora* oocysts demonstrated by the Ocus 40 and EasyScan One scanners were comparable with overlapping 95% CIs (Tables 5, 6). The sensitivity for *Ancylostoma* eggs (80.0%; 95% CI: 49.7–95.6%) recorded with the Ocus 40 scanner was slightly lower than that of other targeted parasites. This result was very likely because only 10 positive fecal samples were available. In addition, the 2 FN samples contained only 2–3 visibly deteriorated *Ancylostoma* eggs on the entire fecal slide. Specificity for *Trichuris* eggs was 100.0% (95% CI: 98.8–100.0%) with the Ocus 40 scanner and 97.7% (95% CI: 95.6–99.0%) with the EasyScan One scanner. We did not assess sensitivities for *Trichuris* eggs in feline fecal samples given the lack of TP samples, which reflects the low prevalence of this infection in cats in most of the United States.<sup>2,6,13,39</sup> However, *Trichuris* infection in cats is frequently reported in St. Kitts, the West Indies,<sup>21</sup> and several different species of *Trichuris* have been described from the Caribbean and Central and South America, including *T. felis*, *T. campanula*, and *T. serrata*.<sup>3,32</sup> Additionally, *Trichuris* has been found in Florida, USA.<sup>13</sup> Further study is required to confirm the capability of detecting and identifying feline *Trichuris* eggs using the Vetscan Imagyst system.

Higher sensitivity for *Giardia* cysts was observed with the Ocus 40 scanner (92.1%; 95% CI: 83.5–96.9%) than with the

**Table 6.** Algorithm performance with the EasyScan One scanner for feline fecal samples prepared with the Vetscan Imagyst sample preparation device with sugar solution: diagnostic sensitivity and specificity comparing the results reported by an expert (reference) versus by the Vetscan Imagyst system.

	Ancylostoma	Cystoisospora	Toxocara	Trichuris
True-positive	27	96	94	0
True-negative	276	189	203	302
False-positive	5	15	7	7
False-negative	1	9	5	0
Total	309	309	309	309
Sensitivity, % (95% CI)	96.4 (84.5–99.6)	91.4 (84.9–95.7)	95.0 (89.3–98.0)	NA
Specificity, % (95% CI)	98.2 (96.1–99.3)	92.7 (88.5–95.6)	96.7 (93.6–98.5)	97.7 (95.6–99.0)

NA=not applicable.

**Table 7.** Algorithm performance with the Ocus 40 and EasyScan One scanners for canine and feline fecal samples prepared with the Vetscan Imagyst sample preparation device with 33% zinc sulfate solution: diagnostic sensitivity and specificity comparing the results reported by an expert (reference) versus by the Vetscan Imagyst system.

	<i>Giardia</i>	
	Ocus 40	EasyScan One
True-positive	58	53
True-negative	84	15
False-positive	1	0
False-negative	5	19
Total	148	87
Sensitivity, % (95% CI)	92.1 (83.5–96.9)	73.6 (62.6–82.7)
Specificity, % (95% CI)	98.8 (94.6–99.9)	100.0 (84.8–100)

EasyScan One scanner (73.6%; 95% CI: 62.6–82.7%). This increase in sensitivity might be the result of the difference in color presentation and production process of images between the 2 scanners and the different compatibility of the algorithm versions applied to the 2 scanners. *Giardia* is a commonly detected gastrointestinal parasite in dogs and cats<sup>12,15,16,18,19,24,28,37,38</sup>; however, *Giardia* is considered one of the more challenging gastrointestinal parasites to detect by fecal examination given its small size, transparency, and intermittent shedding of the cysts and trophozoites.<sup>20,39,40</sup> Additionally, application of an optimal fecal examination method is essential to detect *Giardia* cysts; a centrifugal fecal flotation using 33% zinc sulfate solution is a recommended technique, and recovery of this parasite from fecal samples can be difficult with suboptimal methods, such as a passive flotation technique with sodium nitrate solution.<sup>2,39</sup> Given that *Giardia* infection can be subclinical for many patients, it is important to conduct a fecal examination when new animals are introduced to homes, whether or not animals have clinical signs.<sup>5,9,39</sup> As a supplemental detection test in conjunction with conventional fecal examinations, *Giardia*-specific coproantigen detection assays are commercially

available to facilitate the detection of *Giardia* infection, primarily for clinical patients.<sup>5,9,39</sup> A centrifugal fecal flotation test with 33% zinc sulfate solution is a recommended follow-up test after the completion of treatment because *Giardia* antigens may continue to be excreted in feces after parasite elimination.<sup>5,9,39</sup> The increased sensitivity for detection of *Giardia* cysts by the Vetscan Imagyst system with the Ocus 40 scanner combined with 33% zinc sulfate solution can provide accurate, consistent detection of *Giardia* infection in dogs and cats.

A limitation of the *Giardia* analysis in our study was a relatively low number of TN samples ( $n=15$ ) for the EasyScan One scanner assessment. This limitation could have influenced the DSp for *Giardia* (100.0%; 95% CI: 84.8–100.0%) that was demonstrated by the EasyScan One scanner, although the specificity of *Giardia* observed in the previous study by the same scanner was similarly high (97.0%; 95% CI: 90.8–99.4%).<sup>27</sup>

Using the EasyScan One scanner with the Vetscan Imagyst system, the diagnostic sensitivities for identifying *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, *Cystoisospora* oocysts, and *Giardia* cysts in canine and feline fecal samples were 73.6–96.4% in our present study and 75.8–100.0% in our 2 previous studies.<sup>26,27</sup> Similarly, the diagnostic specificities for *Ancylostoma*, *Toxocara*, and *Trichuris* eggs, and *Giardia* cysts in canine and feline samples aligned among the 3 studies: 79.7–100.0% in our present study and 93.1–100.0% in the previous 2 studies.<sup>26,27</sup> The specificity associated with detection of *Cystoisospora* oocysts in canine samples, however, was lower in our present study (79.7%; 95% CI: 76.8–82.5%) compared to the specificity of 93.1% (95% CI: 88.6–96.2%) in our previous study.<sup>27</sup> This notable difference could have been the result of the different sample size; we tested 943 canine fecal samples in our present study, but only 200 fecal samples (104 canine and 96 feline fecal samples) in our previous study.<sup>27</sup>

Our statistical evaluations were limited by previously defined ground truth values. Given that the manual examination conducted by laboratory personnel was considered ground truth, even when the Vetscan Imagyst system

correctly identified targeted parasites, some results were categorized as FPs because human examiners missed them at the initial manual reading. If these parasites had been detected by the examiners during the initial reading, the sensitivity and specificity values of the Vetscan Imagyst system could have been better than the calculated values in our study. With continuous guided algorithm training, learning, and development, the diagnostic performance of the Vetscan Imagyst system to detect gastrointestinal parasites in feces is expected to improve even further over time.

### Acknowledgments

We thank Emily Looper, Tracey Land, Jamie Butler, and Joy Bowles for assistance in preparation of the fecal samples and examination of the flotation slides. We also thank all of the personnel who were involved in collecting the fecal samples from dogs and cats for our study. Writing and editorial assistance were provided by Litto Communications.

### Declaration of Conflicting Interests

Travis Cree, Michael Loenser, Cory Penn, Austin Rhodes, and Richard Goldstein are employees of Zoetis. Benjamin Larson is an employee of Techcyte. Yoko Nagamori was an employee of Oklahoma State University during our study and is now an employee of Zoetis. Ruth Hall-Sedlak was an employee of Zoetis when our study was conducted and is now an employee of Amgen.

### Funding

Our study was funded by Zoetis.

### ORCID iD

Yoko Nagamori  <https://orcid.org/0000-0002-9515-3215>

### References

- Ballweber LR, et al. American Association of Veterinary Parasitologists' review of veterinary fecal flotation methods and factors influencing their accuracy and use—is there really one best technique? *Vet Parasitol* 2014;204:73–80.
- Bowman DD. Helminths. In: Bowman DD, ed. *Georgis' Parasitology for Veterinarians*. 11th ed. Elsevier, 2019:242.
- Bowman DD. *Trichuris felis*. In: Bowman DD, et al. *Feline Clinical Parasitology*. Iowa State University Press, 2002:348–350.
- Bramão I, et al. The role of color information on object recognition: a review and meta-analysis. *Acta Psychol (Amst)* 2011;138:244–253.
- Companion Animal Parasite Council (CAPC). *Giardia* for dog. 2021. [cited 2021 Nov 10]. <https://capcvet.org/guidelines/giardia/>
- Companion Animal Parasite Council (CAPC). *Trichuris vulpis* for dog. 2021. [cited 2021 Nov 8]. <https://capcvet.org/guidelines/trichuris-vulpis/>
- Dryden MW, et al. Gastrointestinal parasites: the practice guide to accurate diagnosis and treatment. *Compend Contin Educ Vet* 2006;28(Suppl):3–13.
- Elghryani N, et al. Preliminary evaluation of a novel, fully automated, Telenostic device for rapid field-diagnosis of cattle parasites. *Parasitology* 2020;147:1249–1253.
- European Scientific Counsel Companion Animal Parasites (ESCCAP). Control of intestinal protozoa in dogs and cats. In: ESCCAP Guideline 06. 2nd ed. ESCCAP, Feb 2018. [cited 2021 Nov 10]. <http://www.esccap.org/link-document/32/>
- Flachot A, Gegenfurtner KR. Color for object recognition: hue and chroma sensitivity in the deep features of convolutional neural networks. *Vision Res* 2021;182:89–100.
- Flachot A, Gegenfurtner KR. Processing of chromatic information in a deep convolutional neural network. *J Opt Soc Am A Opt Image Sci Vis* 2018;35:B334–B346.
- Gates MC, Nolan TJ. Endoparasite prevalence and recurrence across different age groups of dogs and cats. *Vet Parasitol* 2009;166:153–158.
- Geng J, et al. Diagnosis of feline whipworm infection using a coproantigen ELISA and the prevalence in feral cats in southern Florida. *Vet Parasitol Reg Stud Reports* 2018;14:181–186.
- Grundium Ltd. Grundium Ocus®40. 2022. [cited 2022 Mar 24]. <https://www.grundium.com/ocus40/>
- Hoggard KR, et al. Prevalence survey of gastrointestinal and respiratory parasites of shelter cats in northeastern Georgia, USA. *Vet Parasitol Reg Stud Reports* 2019;16:100270.
- Hoopes JH, et al. A retrospective investigation of feline gastrointestinal parasites in western Canada. *Can Vet J* 2013;54:359–362.
- Inácio SV, et al. Automated diagnosis of canine gastrointestinal parasites using image analysis. *Pathogens* 2020;9:139.
- Jacobs SR, et al. A survey of the prevalence of *Giardia* in dogs presented to Canadian veterinary practices. *Can Vet J* 2001;42:45–46.
- Joffe D, et al. The prevalence of intestinal parasites in dogs and cats in Calgary, Alberta. *Can Vet J* 2011;52:1323–1328.
- Kirkpatrick CE. Giardiasis. *Vet Clin North Am Small Anim Pract* 1987;17:1377–1387.
- Krecek RC, et al. Parasites of stray cats (*Felis domesticus* L., 1758) on St. Kitts, West Indies. *Vet Parasitol* 2010;172:147–149.
- Little SE, et al. Coproantigen detection augments diagnosis of common nematode infections in dogs. *Top Companion Anim Med* 2019;35:42–46.
- Lo F, et al. Pollen calendars and maps of allergenic pollen in North America. *Aerobiologia (Bologna)* 2019;35:613–633.
- Lucio-Forster A, Bowman DD. Prevalence of fecal-borne parasites detected by centrifugal flotation in feline samples from two shelters in upstate New York. *J Feline Med Surg* 2011;13:300–303.
- Motic. MoticEasyScan One. 2023. [cited 2023 Aug 8]. [https://www.motic.com/As\\_MoticEasyScan\\_One1/](https://www.motic.com/As_MoticEasyScan_One1/)
- Nagamori Y, et al. Evaluation of the VETSCAN IMAGYST: an in-clinic canine and feline fecal parasite detection system integrated with a deep learning algorithm. *Parasit Vectors* 2020;13:346.
- Nagamori Y, et al. Further evaluation and validation of the VETSCAN IMAGYST: in-clinic feline and canine fecal parasite detection system integrated with a deep learning algorithm. *Parasit Vectors* 2021;14:89.



28. Olson ME, et al. Prevalence and diagnosis of *Giardia* infection in dogs and cats using a fecal antigen test and fecal smear. *Can Vet J* 2010;51:640–642.
29. Pellikka M, inventor; Grundium Oy, assignee. Method for image stitching. United States patent US 10628698. 2020 Apr 21.
30. Pellikka M, Vartiainen M, inventors; Grundium Oy, assignee. Colour calibration of an imaging device. United States patent US 11006088. 2021 May 11.
31. Rahn JR, et al. Depth of field extension for optical tomography. United States patent application publication US 2009/0103792 A1. 2009 Apr 23.
32. Santa Cruz AM, et al. Parasitological results of 50 necropsies of cats in Corrientes City, Argentina. *Vet Argent* 1987;4:735–739.
33. Scare JA, et al. Evaluation of accuracy and precision of a smartphone based automated parasite egg counting system in comparison to the McMaster and Mini-FLOTAC methods. *Vet Parasitol* 2017;247:85–92.
34. Slusarewicz P, et al. Automated parasite faecal egg counting using fluorescence labelling, smartphone image capture and computational image analysis. *Int J Parasitol* 2016;46:485–493.
35. Sweet S, et al. A 3-year retrospective analysis of canine intestinal parasites: fecal testing positivity by age, U.S. geographical region and reason for veterinary visit. *Parasit Vectors* 2021;14:173.
36. Vartiainen M, et al., investors; Grundium Oy, assignee. Microscope comprising a movable objective-camera system with rolling shutter camera sensor and a multi-color strobe flash and method for scanning microscope slides with proper focus. United States patent US 10564408. 2020 Feb 18.
37. Villeneuve A, et al. Parasite prevalence in fecal samples from shelter dogs and cats across the Canadian provinces. *Parasit Vectors* 2015;8:281.
38. Wyrosdick HM, et al. Parasite prevalence survey in shelter cats in Citrus County, Florida. *Vet Parasitol Reg Stud Reports* 2017;10:20–24.
39. Zajac AM, et al. Fecal examination for the diagnosis of parasitism: dogs and cats. In: Zajac AM, et al. *Veterinary Clinical Parasitology*. 9th ed. Blackwell, 2021:42–96.
40. Zajac AM, et al. Evaluation of the importance of centrifugation as a component of zinc sulfate fecal flotation examinations. *J Am Anim Hosp Assoc* 2002;38:221–224.
41. Zoetis. VETSCAN IMAGYST™. 2021. [cited 2022 Dec 5]. <https://www2.zoetisus.com/products/diagnostics/instruments/vetscan-imagyst>
42. Zoetis. VETSCAN IMAGYST™ Quick Start Guide. 2020. [cited 2022 Dec 5]. [https://www.vetscanimagyst.com/assets/pdf/User-Guide\\_AI-Fecal-Analysis.pdf](https://www.vetscanimagyst.com/assets/pdf/User-Guide_AI-Fecal-Analysis.pdf)